

辐照证明书 CERTIFICATE OF IRRADIATION

No.:CG23110352R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation				2023/11/15			
产品名称Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark			
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)				
0.5ml预制编码冻存管	NA	W0123111545	82	12	——	12.2	——	NA			
NA											
NA											
NA											
剂量测量系统/Dosimetry system WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX_EA 不确定度/Uncertainty: k=2, U=5%											

康盛生物

签发/Issued by:

日期/Date:

2023-11-15

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如客户要求的剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北库社区库西路1288号/ NO.1288, Shexi Road, Beishu, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cgndasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.:C/23120251R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation			2023/12/6		
产品名称Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark	
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)		
0.5ml预制编码冻存管	NA	W0123120648	76	12	—	12.2	—	NA	
NA									
NA									
NA									
WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX_EA 不确定度/Uncertainty: k=2, U=5%									

签发/Issued by:

日期/Date:

2023. 12. 06

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如生物制品）已按照客户要求剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北库社区西西路1288号/ NO.1288, Shexi Road, Beishe, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cgn-dasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.:CI24010156R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation		2024/1/22		
产品名称Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)	
0.5ml预制编码冻存管	NA	W0124012223	62	12	—	12.2	—	NA
NA								
NA								
NA								
剂量测量系统/Dosimetry system	WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX_EC 不确定度/Uncertainty: k=2, U=5%							

检测
/Inspector:

复核
/Reviewer

日期
/Date:

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如生产批次描述），已按照客户要求的剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北库社区库西路1288号/NO.1288, Shexi Road, Beishe, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cindasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.:C124020248R1

客户产品信息/Customer Product Information		客户/Customer 苏州康盛生物有限公司			辐照日期/Date of Irradiation			2024/2/15	
产品名称Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark	
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)		
0.5ml预制编码冻存管	NA	W0124021541	58	12	---	12.2	---	NA	
NA									
NA									
NA									
剂量测量系统/Dosimetry system	WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX 不确定度/Uncertainty: k=2, U=5%								

检测
/Inspector:

复核
/Reviewer

日期
/Date:

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如生产厂商所标注）均按照客户要求剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北厍社区西路1288号/ NO.1288, Shexi Road, Beishe, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cgndasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.: C124030189R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation		2024/3/14		
产品名称/Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)	
0.5ml预制编码冻存管	NA	W0124031434	83	12	---	12.2	---	NA
NA								
NA								
NA								
剂量测量系统/Dosimetry system		WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX_EC 不确定度/Uncertainty: k=2, U=5%						

CEPRON ACCELERATOR TECHNOLOGY

康盛生物

检测 /Inspector:	复核 /Reviewer	日期 /Date:
		

中广核加速器技术有限公司特此证明上述清单中的产品（如生产请见描述），已按照客户要求剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北厍社区西西路1288号/NO.1288, Shexi Road, Beishe, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cngdasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.: C124040378R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation		2024/4/15		
产品名称Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose		备注 Remark
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)	
0.5ml预制编码冻存管	NA	W0124041528	83	12	—	12.2	—	NA
NA								
NA								
NA								
WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX_EC 不确定度/Uncertainty: k=2, U=5%								

检测
/Inspector:

复核
/Reviewer

日期
/Date:

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如生产商所描述），已按照客户要求的剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.

地址/Address：中国江苏省苏州市吴江区黎里镇北厍社区库西路1288号/ NO.1288, Shexi Road, Beishe, Wujiang, Suzhou, Jiangsu, China
电话/Tel：086-512-82859888-8879 <http://www.cndasheng.com>

辐照证明书 CERTIFICATE OF IRRADIATION

No.:C124050321R1

客户产品信息/Customer Product Information		客户/Customer:苏州康盛生物有限公司		辐照日期/Date of Irradiation			2024/5/9		备注 Remark
产品名称/Products	批号 Lot No.	辐照批号 Irradiation Batch	数量QTY (箱CS)	要求剂量 Specified Dose		实际剂量 Actual Dose			
				Min Dose(kGy)	Max Dose(kGy)	Min Dose(kGy)	Max Dose(kGy)		
0.5ml预制编码冻存管	NA	W0124050931	49	12	——	12.2	——	NA	
NA									
NA									
NA									
WINDose 剂量测量系统/WINDose Dosimetry system 剂量计批号/Dosimeter Lot No.: GEX-ACCELERATOR TECHNOLOGY 不确定度/Uncertainty: k=2, U=5%									

检测
/Inspector:

复核
/Reviewer

日期
Date:

2024.5.9

中广核康盛生物

辐照部

中广核达胜加速器技术有限公司特此证明上述清单中的产品（如生产商所描述），已按照客户要求剂量进行电子束辐照，产品实际剂量如上所述。
CGN DASHENG ELECTRON ACCELERATOR TECHNOLOGY CO.,LTD certifies that the products listed above (as described by the Manufacturer) have been irradiated by electron beam according to the dose specified by the customer, and the actual dose is as described above.